

Quality Control Certificate

Product: Standard Column

Product No.: 13375

Lot No.: 1000171

Storage Recommendations: Store the column at room temperature below 25°C

Description: The standard column is part of a 3- or 4-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned-up on a DEXTech Plus system, spiked with recovery standard, evaporated via DEva and has been quantified with a HRGC/HRMS with a resolution of R > 10000.

Results Blank Value:

| | | |
|--------------------|----------|----------------|
| PCDD/F-TEQ: | 0,2 | pg/column |
| | (crit: < | 0,7 pg/column |
| dl-PCB-TEQ: | 0,01 | pg/column |
| | (crit: < | 0,05 pg/column |
| Sum Indikator PCB: | 18,02 | pg/column |
| | (crit: < | 100 pg/column |

Results Recoveries:

| | | | | | |
|--------|----|----|-----|---|--------------------|
| PCDD/F | 98 | to | 114 | % | (crit: 70 to 120) |
| PCB | 91 | to | 98 | % | (crit: 70 to 120) |

This is to certify that standard column, Lot 1000171, passed the required test specifications and is released for sale.

date: 08.02.2021 sign.: _____

T. Kehmeier

The company LCTech GmbH is certified according to ISO 9001:2015



Hazards: NOT FOR HUMAN OR DRUG USE!

The standard column is designed and prepared for usage with the alumina/florisil column and carbon column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.

Quality Control: All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.

Quality Management: This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)

Documentation / Data Attached: Table 1 & 2: Blank values of PCDD/F and PCB
Table 3 & 4: 13C-Recoveries of PCDD/F and PCB

Analytcs: All the columns (n>5) have to perform a clean-up of a solvent blank (10 mL n-hexane), spiked with a 13C - labelled quantifier-standard solution with a single column method onto a DEXTech Plus system. The fractions 1 (PCB) and 2 (PCDD/F) are spiked with 13C - labelled recovery- standard solutions and evaporated with the D-EVA vacuum centrifuge. The extracts are measured with a HRMS-DFS from Thermo Fisher Scientific with a resolution of R > 10000. The HRGCs are equipped with 60 m DB5 MS columns. For PCDD/F 5µL are injected via PTV, for PCB 2µL via SSL.

Remarks: Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of silica or filters without any effect on the clean-up.

The company LCTech GmbH is certified according to ISO 9001:2015

Results:

Lockmass check:

No significant disturbances, or indicators for contaminations are detected.

Blanks:

Table 1: PCDD/F blank (n=6)

| Congeneres: | [pg/column]: |
|---------------------|--------------|
| 2,3,7,8-TCDF | 0,07 |
| 1,2,3,7,8-PeCDF | 0,13 |
| 2,3,4,7,8-PeCDF | 0,1 |
| 1,2,3,4,7,8-HxCDF | 0,057 |
| 1,2,3,6,7,8-HxCDF | 0,067 |
| 2,3,4,6,7,8-HxCDF | 0,06 |
| 1,2,3,7,8,9-HxCDF | 0,09 |
| 1,2,3,4,6,7,8-HpCDF | <0,063 |
| 1,2,3,4,7,8,9-HpCDF | 0,02 |
| OCDF | 0,08 |
| 2,3,7,8-TCDD | <dl |
| 1,2,3,7,8-PeCDD | 0,1 |
| 1,2,3,4,7,8-HxCDD | 0,055 |
| 1,2,3,6,7,8-HxCDD | 0,12 |
| 1,2,3,7,8,9-HxCDD | 0,09 |
| 1,2,3,4,6,7,8-HpCDD | 0,1 |
| OCDD | 0,34 |

| TEQ (WHO 2005) | |
|----------------|------|
| lower bound | 0,2 |
| upper bound | 0,21 |

Table 2: PCB blank (n=6)

| Congeneres: | [pg/column]: |
|-------------|--------------|
| PCB 28 | 3,3 |
| PCB 52 | 4,59 |
| PCB 77 | 0,18 |
| PCB 81 | 0,056 |
| PCB 101 | 3,53 |
| PCB 123 | 0,2633 |
| PCB 118 | 1,51 |
| PCB 114 | 0,1372 |
| PCB 105 | 0,64 |
| PCB 126 | 0,0912 |
| PCB 153 | 2,79 |
| PCB 138 | 2,67 |
| PCB 167 | 0,544 |
| PCB 156 | 0,26 |
| PCB 157 | 0,3 |
| PCB 169 | <0,027 |
| PCB 180 | 0,6 |
| PCB 189 | 0,217 |

| TEQ (WHO 2005) | |
|----------------|--------|
| lower bound | 0,0097 |
| upper bound | 0,0097 |

| | |
|-------------|-------|
| Sum DIN PCB | 18,02 |
|-------------|-------|

The company LCTech GmbH is certified according to ISO 9001:2015



Results:

13C-Recoveries

Table 3: PCDD/F 13C-recoveries (n=6)

| Congeneres: | 13C rec [%] |
|---------------------|-------------|
| 2,3,7,8-TCDF | 98 |
| 1,2,3,7,8-PeCDF | 105 |
| 2,3,4,7,8-PeCDF | 101 |
| 1,2,3,4,7,8-HxCDF | 108 |
| 1,2,3,6,7,8-HxCDF | 106 |
| 2,3,4,6,7,8-HxCDF | 104 |
| 1,2,3,7,8,9-HxCDF | 109 |
| 1,2,3,4,6,7,8-HpCDF | 114 |
| 1,2,3,4,7,8,9-HpCDF | 113 |
| OCDF | 108 |
| 2,3,7,8-TCDD | 104 |
| 1,2,3,7,8-PeCDD | 107 |
| 1,2,3,4,7,8-HxCDD | 111 |
| 1,2,3,6,7,8-HxCDD | 111 |
| 1,2,3,7,8,9-HxCDD | 107 |
| 1,2,3,4,6,7,8-HpCDD | 107 |
| OCDD | 110 |
| | |

Table 4: PCB 13C-recoveries (n=6)

| Congeneres: | 13C rec [%] |
|-------------|-------------|
| PCB 28 | 93 |
| PCB 52 | 93 |
| PCB 77 | 98 |
| PCB 81 | 91 |
| PCB 101 | 96 |
| PCB 123 | 98 |
| PCB 118 | 96 |
| PCB 114 | 97 |
| PCB 105 | 94 |
| PCB 126 | 94 |
| PCB 153 | 94 |
| PCB 138 | 92 |
| PCB 167 | 95 |
| PCB 156 | 93 |
| PCB 157 | 94 |
| PCB 169 | 97 |
| PCB 180 | 97 |
| PCB 189 | 95 |

The company LCTech GmbH is certified according to ISO 9001:2015

