

# Quality Control Certificate

**Product:** Alumina Column

**Product No.:** 15433

**Lot No.:** 714093

**Storage Recommendations:** Store the column at room temperature below 25°C

---

---

**Description:** The alumina column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

## Quality Control Release Inspection and Test Specification

**Test Procedure:** A solvent blank, spiked with quantification standard has been cleaned-up on a DEXTech Plus system, spiked with recovery standard, evaporated via DEva and has been quantified with a HRGC/HRMS with a resolution of R > 10000.

**Results Blank Value:**

PCDD/F-TEQ:	0,06	pg/column
	(crit: <	0,7 pg/column
dl-PCB-TEQ:	0,012	pg/column
	(crit: <	0,05 pg/column
Sum Indikator PCB:	7,2	pg/column
	(crit: <	100 pg/column

**Results Recoveries:**

PCDD/F	70	to	119	%	(crit: 70 to 120 )
PCB	70	to	92	%	(crit: 70 to 120 )

---

---

This is to certify that alumina column, Lot 714093, passed the required test specifications and is released for sale.

date: 10.12.2020 sign.: \_\_\_\_\_

*T. Kehmeier*

The company LCTech GmbH is certified according to ISO 9001:2015



**Hazards:** NOT FOR HUMAN OR DRUG USE!

The alumina column is designed and prepared for usage with the universal/standard & smart column and carbon column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.

**Quality Control:** All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.

**Quality Management:** This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)

**Documentation / Data Attached:** Table 1 & 2: Blank values of PCDD/F and PCB  
Table 3 & 4: 13C-Recoveries of PCDD/F and PCB

**Analytics:** All the columns (n>5) have to perform a clean-up of a solvent blank (10 mL n-hexane), spiked with a 13C - labelled quantifier-standard solution with a single column method onto a DEXTech Plus system. The fractions 1 (PCB) and 2 (PCDD/F) are spiked with 13C - labelled recovery- standard solutions and evaporated with the D-EVA vacuum centrifuge. The extracts are measured with a HRMS-DFS from Thermo Fisher Scientific with a resolution of R > 10000. The HRGCs are equipped with 60 m DB5 MS columns. For PCDD/F 5µL are injected via PTV, for PCB 2µL via SSL.

**Remarks:** n/a

*The company LCTech GmbH is certified according to ISO 9001:2015*



**Results:**

Lockmass check:

No significant disturbances, or indicators for contaminations are detected.

Blanks:

Table 1: PCDD/F blank (n=7)

Congeneres:	[pg/column]:
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	0,05
2,3,4,7,8-PeCDF	<0,081
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	<dl
2,3,4,6,7,8-HxCDF	<dl
1,2,3,7,8,9-HxCDF	0,05
1,2,3,4,6,7,8-HpCDF	<dl
1,2,3,4,7,8,9-HpCDF	0,196
OCDF	1,86
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	<0,054
1,2,3,4,7,8-HxCDD	<dl
1,2,3,6,7,8-HxCDD	<dl
1,2,3,7,8,9-HxCDD	<dl
1,2,3,4,6,7,8-HpCDD	<dl
OCDD	<dl

TEQ (WHO 2005)	
lower bound	0,06
upper bound	0,08

Table 2: PCB blank (n=7)

Congeneres:	[pg/column]:
PCB 28	1,2
PCB 52	2,93
PCB 77	0,13
PCB 81	0,172
PCB 101	1,49
PCB 123	0,0313
PCB 118	0,41
PCB 114	0,4858
PCB 105	0,16
PCB 126	0,0977
PCB 153	0,62
PCB 138	0,56
PCB 167	0,055
PCB 156	0,15
PCB 157	0,107
PCB 169	0,055
PCB 180	0,36
PCB 189	0,329

TEQ (WHO 2005)	
lower bound	0,0116
upper bound	0,0116

Sum DIN PCB	7,2
-------------	-----

The company LCTech GmbH is certified according to ISO 9001:2015

**Results:**

13C-Recoveries

Table 3: PCDD/F 13C-recoveries (n=7)

Congeneres:	13C rec [%]
2,3,7,8-TCDF	103
1,2,3,7,8-PeCDF	106
2,3,4,7,8-PeCDF	99
1,2,3,4,7,8-HxCDF	86
1,2,3,6,7,8-HxCDF	91
2,3,4,6,7,8-HxCDF	86
1,2,3,7,8,9-HxCDF	70
1,2,3,4,6,7,8-HpCDF	119
1,2,3,4,7,8,9-HpCDF	114
OCDF	92
2,3,7,8-TCDD	102
1,2,3,7,8-PeCDD	108
1,2,3,4,7,8-HxCDD	89
1,2,3,6,7,8-HxCDD	96
1,2,3,7,8,9-HxCDD	90
1,2,3,4,6,7,8-HpCDD	119
OCDD	83

Table 4: PCB 13C-recoveries (n=7)

Congeneres:	13C rec [%]
PCB 28	83
PCB 52	72
PCB 77	81
PCB 81	78
PCB 101	75
PCB 123	92
PCB 118	89
PCB 114	85
PCB 105	88
PCB 126	84
PCB 153	72
PCB 138	77
PCB 167	78
PCB 156	79
PCB 157	81
PCB 169	76
PCB 180	73
PCB 189	70

The company LCTech GmbH is certified according to ISO 9001:2015

