

# Quality Control Certificate

**Product:** Alumina Column

**Product No.:** 15433

**Lot No.:** 716447

**Storage Recommendations:** Store the column at room temperature below 25°C

**Description:** The alumina column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

## Quality Control Release Inspection and Test Specification

**Test Procedure:** A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated via DEva and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

**Results Blank Value:**

PCDD/F-TEQ:	0,02	pg/column
	(crit: <	0,7 pg/column)
dl-PCB-TEQ:	0,019	pg/column
	(crit: <	0,05 pg/column)
Sum Indikator PCB:	4,8	pg/column
	(crit: <	100 pg/column)

**Results Recoveries:**

PCDD/F	72	to	105	%	(crit: 70	to	120)
PCB	86	to	109	%	(crit: 70	to	120)

This is to certify that the alumina column, Lot 716447, passed the required test specifications and is released for sale.

date: 03.03.2022 sign.: \_\_\_\_\_

*T. Kehmeier*

The company LCTech GmbH is certified according to ISO 9001:2015



**Hazards:** NOT FOR HUMAN OR DRUG USE!

The alumina column is designed and prepared for usage with the universal/standard & smart column and carbon column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.

**Quality Control:** All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.

**Quality Management:** This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)

**Documentation / Data Attached:** Table 1 & 2: Blank values of PCDD/F and PCB  
Table 3 & 4: 13C-Recoveries of PCDD/F and PCB

**Analytcs:** All the columns (n>5) have to perform a clean-up of a solvent blank (10 mL n-hexane), spiked with a 13C - labelled quantifier-standard solution with a single column method onto a DEXTech Plus system. The fractions 1 (PCB) and 2 (PCDD/F) are spiked with 13C - labelled recovery- standard solutions and evaporated with the D-EVA vacuum centrifuge. The extracts are measured with a HRMS-DFS from Thermo Fisher Scientific with a resolution of R > 10000. The HRGCs are equipped with 60 m DB5 MS columns. For PCDD/F 5µL are injected via PTV, for PCB 2µL via SSL.

**Remarks:** n/a

*The company LCTech GmbH is certified according to ISO 9001:2015*



**Results:**

Lockmass check:

No significant disturbances, or indicators for contaminations are detected.

Blanks:

Table 1: PCDD/F blank (n=5)

Congeneres:	[pg/column]:
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	<dl
2,3,4,6,7,8-HxCDF	<dl
1,2,3,7,8,9-HxCDF	<dl
1,2,3,4,6,7,8-HpCDF	<dl
1,2,3,4,7,8,9-HpCDF	0,064
OCDF	<0,054
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	<0,054
1,2,3,4,7,8-HxCDD	<dl
1,2,3,6,7,8-HxCDD	<dl
1,2,3,7,8,9-HxCDD	<dl
1,2,3,4,6,7,8-HpCDD	<dl
OCDD	0,19

TEQ (WHO 2005)	
lower bound	0,02
upper bound	0,05

Table 2: PCB blank (n=5)

Congeneres:	[pg/column]:
PCB 28	1,11
PCB 52	1,51
PCB 77	0,79
PCB 81	0,147
PCB 101	1,04
PCB 123	0,0326
PCB 118	0,24
PCB 114	0,0522
PCB 105	0,09
PCB 126	0,1362
PCB 153	0,26
PCB 138	<dl
PCB 167	0,133
PCB 156	0,44
PCB 157	0,234
PCB 169	0,17
PCB 180	0,67
PCB 189	0,148

TEQ (WHO 2005)	
lower bound	0,0189
upper bound	0,0189

Sum DIN PCB	4,8
-------------	-----

The company LCTech GmbH is certified according to ISO 9001:2015

**Results:**

13C-Recoveries

Table 3: PCDD/F 13C-recoveries (n=5)

Congeneres:	13C rec [%]
2,3,7,8-TCDF	73
1,2,3,7,8-PeCDF	81
2,3,4,7,8-PeCDF	78
1,2,3,4,7,8-HxCDF	72
1,2,3,6,7,8-HxCDF	80
2,3,4,6,7,8-HxCDF	81
1,2,3,7,8,9-HxCDF	89
1,2,3,4,6,7,8-HpCDF	85
1,2,3,4,7,8,9-HpCDF	93
OCDF	105
2,3,7,8-TCDD	76
1,2,3,7,8-PeCDD	82
1,2,3,4,7,8-HxCDD	80
1,2,3,6,7,8-HxCDD	72
1,2,3,7,8,9-HxCDD	88
1,2,3,4,6,7,8-HpCDD	83
OCDD	96

Table 4: PCB 13C-recoveries (n=5)

Congeneres:	13C rec [%]
PCB 28	98
PCB 52	109
PCB 77	106
PCB 81	99
PCB 101	99
PCB 123	95
PCB 118	90
PCB 114	105
PCB 105	90
PCB 126	86
PCB 153	95
PCB 138	94
PCB 167	98
PCB 156	91
PCB 157	88
PCB 169	88
PCB 180	90
PCB 189	89

The company LCTech GmbH is certified according to ISO 9001:2015

