

## Quality Control Certificate

Product: **Alumina Column**  
Product No.: 15433  
Lot No.: **719527**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

### Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:

PCDD/F-TEQ:	0,18	pg/column
	(crit: <	0,7 pg/column)
dl-PCB-TEQ:	0,0355	pg/column
	(crit: <	0,05 pg/column)
Sum Total PCB:	20,8	pg/column
	(crit: <	300 pg/column)

Results Recoveries:

PCDD/F	83	to	114	%	(crit: 70	to	120	%)
PCB	70	to	97	%	(crit: 70	to	120	%)

This is to certify that the Alumina Column, Lot 719527, passed the required test specifications and is released for sale.

date: 05.12.2023

sign.: 

The company LCTech GmbH is certified according to ISO 9001



## QC-Certificate - 15433 - 719527

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Alumina Column is designed and prepared for usage with the Universal/standard &amp; Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 &amp; 2: blankvalues of PCDD/F and PCB table 3 &amp; 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Alumina Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>n/a</p>



QC-Certificate - 15433 - 719527

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	0,18
2,3,4,7,8-PeCDF	0,12
1,2,3,4,7,8-HxCDF	0,038
1,2,3,6,7,8-HxCDF	0,037
2,3,4,6,7,8-HxCDF	<0,045
1,2,3,7,8,9-HxCDF	0,07
1,2,3,4,6,7,8-HpCDF	<0,063
1,2,3,4,7,8,9-HpCDF	<0,018
1,2,3,4,6,7,8,9-OCDF	<0,054
2,3,7,8-TCDD	0,05
1,2,3,7,8-PeCDD	<0,054
1,2,3,4,7,8-HxCDD	0,041
1,2,3,6,7,8-HxCDD	0,11
1,2,3,7,8,9-HxCDD	0,072
1,2,3,4,6,7,8-HpCDD	0,09
1,2,3,4,6,7,8,9-OCDD	0,58

Table 2: PCB blank

	[pg/column]
PCB-#28	3,62
PCB-#52	5,51
PCB-#101	3,78
PCB-#153	3,07
PCB-#138	3,3
PCB-#180	1,492
PCB-#81	0,06
PCB-#77	0,28
PCB-#126	0,3306
PCB-#169	0,074
PCB-#123	0,41
PCB-#118	1,54
PCB-#114	0,442
PCB-#105	0,89
PCB-#167	0,285
PCB-#156	0,445
PCB-#157	0,24
PCB-#189	0,582

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,18
upper bound	0,18

PCB-TEQ	[pg/column]
lower bound	0,0355
upper bound	0,0355
Sum DIN	20,8

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	92	6
	1,2,3,7,8-PeCDF	83	9
	2,3,4,7,8-PeCDF	89	12
	1,2,3,4,7,8-HxCDF	105	6
	1,2,3,6,7,8-HxCDF	114	4
	2,3,4,6,7,8-HxCDF	104	5
	1,2,3,7,8,9-HxCDF	101	4
	1,2,3,4,6,7,8-HpCDF	103	4
	1,2,3,4,7,8,9-HpCDF	96	8
	1,2,3,4,6,7,8,9-OCDF	96	3
	2,3,7,8-TCDD	87	8
	1,2,3,7,8-PeCDD	89	9
	1,2,3,4,7,8-HxCDD	109	6
	1,2,3,6,7,8-HxCDD	91	5
	1,2,3,7,8,9-HxCDD	104	5
	1,2,3,4,6,7,8-HpCDD	96	5
	1,2,3,4,6,7,8,9-OCDD	92	3

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	87	5
	PCB-#52	84	6
	PCB-#101	93	7
	PCB-#153	97	6
	PCB-#138	92	5
	PCB-#180	96	9
	PCB-#81	77	10
	PCB-#77	81	9
	PCB-#126	73	12
	PCB-#169	70	15
	PCB-#123	85	8
	PCB-#118	80	9
	PCB-#114	89	8
	PCB-#105	84	7
	PCB-#167	86	10
	PCB-#156	91	8
	PCB-#157	92	9
	PCB-#189	88	10