

Quality Control Certificate

Product: **Alumina Column**
Product No.: 15433
Lot No.: **719663**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:

PCDD/F-TEQ:	0,44	pg/column
	(crit: <	0,7 pg/column)
dl-PCB-TEQ:	0,0255	pg/column
	(crit: <	0,05 pg/column)
Sum Total PCB:	22,5	pg/column
	(crit: <	300 pg/column)

Results Recoveries:

PCDD/F	78	to	94	%	(crit: 70	to	120	%)
PCB	76	to	86	%	(crit: 70	to	120	%)

This is to certify that the Alumina Column, Lot 719663, passed the required test specifications and is released for sale.

date: 26.01.2024

sign.: T. Keshmeir

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 15433 - 719663

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Alumina Column is designed and prepared for usage with the Universal/standard & Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Alumina Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>n/a</p>



QC-Certificate - 15433 - 719663

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	0,04
1,2,3,7,8-PeCDF	0,19
2,3,4,7,8-PeCDF	0,24
1,2,3,4,7,8-HxCDF	0,1
1,2,3,6,7,8-HxCDF	0,152
2,3,4,6,7,8-HxCDF	0,2
1,2,3,7,8,9-HxCDF	0,21
1,2,3,4,6,7,8-HpCDF	0,16
1,2,3,4,7,8,9-HpCDF	0,146
1,2,3,4,6,7,8,9-OCDF	0,17
2,3,7,8-TCDD	0,04
1,2,3,7,8-PeCDD	0,17
1,2,3,4,7,8-HxCDD	0,232
1,2,3,6,7,8-HxCDD	0,34
1,2,3,7,8,9-HxCDD	0,182
1,2,3,4,6,7,8-HpCDD	0,23
1,2,3,4,6,7,8,9-OCDD	0,76

Table 2: PCB blank

	[pg/column]
PCB-#28	2,68
PCB-#52	4,67
PCB-#101	5,33
PCB-#153	4,11
PCB-#138	3,24
PCB-#180	2,425
PCB-#81	0,22
PCB-#77	0,605
PCB-#126	0,1268
PCB-#169	0,415
PCB-#123	0,4
PCB-#118	1,49
PCB-#114	0,436
PCB-#105	1
PCB-#167	0,722
PCB-#156	0,61
PCB-#157	0,66
PCB-#189	0,968

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,44
upper bound	0,44

PCB-TEQ	[pg/column]
lower bound	0,0255
upper bound	0,0255
Sum DIN	22,5

QC-Certificate - 15433 - 719663

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	91	7
	1,2,3,7,8-PeCDF	83	10
	2,3,4,7,8-PeCDF	88	8
	1,2,3,4,7,8-HxCDF	83	6
	1,2,3,6,7,8-HxCDF	94	6
	2,3,4,6,7,8-HxCDF	93	9
	1,2,3,7,8,9-HxCDF	92	7
	1,2,3,4,6,7,8-HpCDF	92	3
	1,2,3,4,7,8,9-HpCDF	84	5
	1,2,3,4,6,7,8,9-OCDF	92	10
	2,3,7,8-TCDD	82	9
	1,2,3,7,8-PeCDD	90	7
	1,2,3,4,7,8-HxCDD	94	9
	1,2,3,6,7,8-HxCDD	78	8
	1,2,3,7,8,9-HxCDD	94	5
	1,2,3,4,6,7,8-HpCDD	86	5
	1,2,3,4,6,7,8,9-OCDD	86	11

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	79	7
	PCB-#52	81	5
	PCB-#101	82	5
	PCB-#153	86	3
	PCB-#138	83	4
	PCB-#180	82	3
	PCB-#81	78	7
	PCB-#77	84	11
	PCB-#126	81	13
	PCB-#169	85	18
	PCB-#123	77	4
	PCB-#118	76	6
	PCB-#114	81	5
	PCB-#105	76	5
	PCB-#167	81	4
	PCB-#156	78	8
	PCB-#157	78	8
	PCB-#189	77	7