

Quality Control Certificate

Product: **Alumina Column**
Product No.: 15433
Lot No.: **721844**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,3	pg/column						
		(crit: <	0,70	pg/column)					
	dl-PCB-TEQ:	0,0397	pg/column						
		(crit: <	0,05	pg/column)					
	Sum Total PCB:	4,6	pg/column						
		(crit: <	300	pg/column)					
Results Recoveries:	PCDD/F	86	to	119	%	(crit: 70	to	120	%)
	PCB	90	to	100	%	(crit: 70	to	120	%)

This is to certify that the Alumina Column, Lot 721844, passed the required test specifications and is released for sale.

date: 07.05.2025

sign.: 

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 15433 - 721844

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Alumina Column is designed and prepared for usage with the Universal/standard & Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Alumina Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>n/a</p>

QC-Certificate - 15433 - 721844

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	0,05
1,2,3,7,8-PeCDF	0,12
2,3,4,7,8-PeCDF	0,1
1,2,3,4,7,8-HxCDF	0,065
1,2,3,6,7,8-HxCDF	0,053
2,3,4,6,7,8-HxCDF	0,05
1,2,3,7,8,9-HxCDF	0,12
1,2,3,4,6,7,8-HpCDF	0,18
1,2,3,4,7,8,9-HpCDF	0,098
1,2,3,4,6,7,8,9-OCDF	0,15
2,3,7,8-TCDD	0,07
1,2,3,7,8-PeCDD	0,1
1,2,3,4,7,8-HxCDD	0,137
1,2,3,6,7,8-HxCDD	0,28
1,2,3,7,8,9-HxCDD	0,127
1,2,3,4,6,7,8-HpCDD	0,27
1,2,3,4,6,7,8,9-OCDD	2,44

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,3
upper bound	0,3

Table 2: PCB blank

	[pg/column]
PCB-#28	1,45
PCB-#52	1,43
PCB-#101	0,52
PCB-#153	0,46
PCB-#138	0,39
PCB-#180	0,313
PCB-#81	0,25
PCB-#77	0,345
PCB-#126	0,24
PCB-#169	0,515
PCB-#123	0,45
PCB-#118	0,43
PCB-#114	0,004
PCB-#105	<0,081
PCB-#167	0,441
PCB-#156	0,272
PCB-#157	0,46
PCB-#189	0,25

PCB-TEQ	[pg/column]
lower bound	0,0397
upper bound	0,0397
Sum DIN	4,6

QC-Certificate - 15433 - 721844

Table 3: PCDD/F recoveries

		[%]	RSD [%]
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	91	3
	1,2,3,7,8-PeCDF	94	4
	2,3,4,7,8-PeCDF	95	3
	1,2,3,4,7,8-HxCDF	103	6
	1,2,3,6,7,8-HxCDF	114	5
	2,3,4,6,7,8-HxCDF	116	5
	1,2,3,7,8,9-HxCDF	104	5
	1,2,3,4,6,7,8-HpCDF	109	5
	1,2,3,4,7,8,9-HpCDF	98	5
	1,2,3,4,6,7,8,9-OCDF	100	4
	2,3,7,8-TCDD	86	3
	1,2,3,7,8-PeCDD	103	2
	1,2,3,4,7,8-HxCDD	115	5
	1,2,3,6,7,8-HxCDD	99	5
	1,2,3,7,8,9-HxCDD	119	4
	1,2,3,4,6,7,8-HpCDD	107	4
	1,2,3,4,6,7,8,9-OCDD	101	4

Table 4: PCB recoveries

		[%]	RSD [%]
PCB 13C Recoveries [%]	PCB-#28	95	3
	PCB-#52	94	4
	PCB-#101	96	2
	PCB-#153	100	4
	PCB-#138	100	0
	PCB-#180	96	4
	PCB-#81	98	1
	PCB-#77	96	4
	PCB-#126	97	7
	PCB-#169	90	3
	PCB-#123	100	5
	PCB-#118	98	5
	PCB-#114	95	6
	PCB-#105	98	7
	PCB-#167	95	2
	PCB-#156	97	4
	PCB-#157	94	4
	PCB-#189	99	2