

Quality Control Certificate

Product: **Alumina Column**
 Product No.: 15433
 Lot No.: **723410**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,23	pg/column
		(crit: <	0,70 pg/column)
	dl-PCB-TEQ:	0,0164	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	27,8	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	87	to	110	%	(crit: 70	to	120	%)
	PCB	71	to	110	%	(crit: 70	to	120	%)

This is to certify that the Alumina Column, Lot 723410, passed the required test specifications and is released for sale.

date: 06.03.2026 sign.: T. Kehmeier

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 15433 - 723410

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Alumina Column is designed and prepared for usage with the Universal/standard & Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Alumina Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>n/a</p>



QC-Certificate - 15433 - 723410

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	0,05
1,2,3,7,8-PeCDF	0,07
2,3,4,7,8-PeCDF	<0,081
1,2,3,4,7,8-HxCDF	<0,027
1,2,3,6,7,8-HxCDF	0,023
2,3,4,6,7,8-HxCDF	0,09
1,2,3,7,8,9-HxCDF	<0,045
1,2,3,4,6,7,8-HpCDF	<0,063
1,2,3,4,7,8,9-HpCDF	0,023
1,2,3,4,6,7,8,9-OCDF	0,23
2,3,7,8-TCDD	<0,036
1,2,3,7,8-PeCDD	0,14
1,2,3,4,7,8-HxCDD	0,042
1,2,3,6,7,8-HxCDD	<0,108
1,2,3,7,8,9-HxCDD	0,035
1,2,3,4,6,7,8-HpCDD	0,1
1,2,3,4,6,7,8,9-OCDD	0,29

Table 2: PCB blank

	[pg/column]
PCB-#28	12,97
PCB-#52	10,15
PCB-#101	1,59
PCB-#153	1,1
PCB-#138	1,19
PCB-#180	0,762
PCB-#81	0,12
PCB-#77	0,105
PCB-#126	0,14
PCB-#169	0,078
PCB-#123	0,06
PCB-#118	0,7
PCB-#114	0,032
PCB-#105	0,22
PCB-#167	0,068
PCB-#156	<0,126
PCB-#157	0,08
PCB-#189	0,132

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,23
upper bound	0,23

PCB-TEQ	[pg/column]
lower bound	0,0164
upper bound	0,0164
Sum DIN	27,8

Table 3: PCDD/F recoveries

	[%]	RSD [%]
2,3,7,8-TCDF	91	7
1,2,3,7,8-PeCDF	105	16
2,3,4,7,8-PeCDF	102	11
1,2,3,4,7,8-HxCDF	98	4
1,2,3,6,7,8-HxCDF	104	5
2,3,4,6,7,8-HxCDF	105	7
1,2,3,7,8,9-HxCDF	106	6
1,2,3,4,6,7,8-HpCDF	101	2
1,2,3,4,7,8,9-HpCDF	92	4
1,2,3,4,6,7,8,9-OCDF	106	5
2,3,7,8-TCDD	90	2
1,2,3,7,8-PeCDD	104	6
1,2,3,4,7,8-HxCDD	107	5
1,2,3,6,7,8-HxCDD	87	3
1,2,3,7,8,9-HxCDD	110	4
1,2,3,4,6,7,8-HpCDD	94	2
1,2,3,4,6,7,8,9-OCDD	96	4

Table 4: PCB recoveries

	[%]	RSD [%]
PCB-#28	97	5
PCB-#52	110	8
PCB-#101	101	3
PCB-#153	88	3
PCB-#138	92	3
PCB-#180	93	5
PCB-#81	76	2
PCB-#77	76	3
PCB-#126	97	10
PCB-#169	82	4
PCB-#123	71	9
PCB-#118	81	5
PCB-#114	81	5
PCB-#105	91	7
PCB-#167	93	2
PCB-#156	91	8
PCB-#157	88	8
PCB-#189	98	3