

Quality Control Certificate

Product: **Alumina Column**
 Product No.: 15433
 Lot No.: **723917**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,09	pg/column
		(crit: <	0,70 pg/column)
	dl-PCB-TEQ:	0,0085	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	9,1	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	89	to	110	%	(crit: 70	to	120	%)
	PCB	86	to	110	%	(crit: 70	to	120	%)

This is to certify that the Alumina Column, Lot 723917, passed the required test specifications and is released for sale.

date: 11.06.2026 sign.: *M. Brack's*

The company LCTech GmbH is certified according to ISO 9001

QC-Certificate - 15433 - 723917

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Alumina Column is designed and prepared for usage with the Universal/standard & Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Alumina Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>n/a</p>

QC-Certificate - 15433 - 723917
Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<0,036
1,2,3,7,8-PeCDF	<0,045
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<0,027
1,2,3,6,7,8-HxCDF	<0,018
2,3,4,6,7,8-HxCDF	<0,045
1,2,3,7,8,9-HxCDF	<0,045
1,2,3,4,6,7,8-HpCDF	<0,063
1,2,3,4,7,8,9-HpCDF	<0,018
1,2,3,4,6,7,8,9-OCDF	0,8
2,3,7,8-TCDD	<0,036
1,2,3,7,8-PeCDD	<0,054
1,2,3,4,7,8-HxCDD	<0,027
1,2,3,6,7,8-HxCDD	<0,108
1,2,3,7,8,9-HxCDD	<0,027
1,2,3,4,6,7,8-HpCDD	0,14
1,2,3,4,6,7,8,9-OCDD	0,87

Table 2: PCB blank

	[pg/column]
PCB-#28	3,27
PCB-#52	2,8
PCB-#101	1,1
PCB-#153	0,72
PCB-#138	0,78
PCB-#180	0,413
PCB-#81	0,11
PCB-#77	0,245
PCB-#126	0,06
PCB-#169	0,081
PCB-#123	0,07
PCB-#118	0,4
PCB-#114	0,055
PCB-#105	<0,081
PCB-#167	0,174
PCB-#156	<0,126
PCB-#157	0,06
PCB-#189	0,091

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,08
upper bound	0,09

PCB-TEQ	[pg/column]
lower bound	0,0085
upper bound	0,0085
Sum DIN	9,1

QC-Certificate - 15433 - 723917
Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	93	4
	1,2,3,7,8-PeCDF	100	5
	2,3,4,7,8-PeCDF	102	5
	1,2,3,4,7,8-HxCDF	97	7
	1,2,3,6,7,8-HxCDF	105	5
	2,3,4,6,7,8-HxCDF	104	5
	1,2,3,7,8,9-HxCDF	110	8
	1,2,3,4,6,7,8-HpCDF	110	3
	1,2,3,4,7,8,9-HpCDF	95	4
	1,2,3,4,6,7,8,9-OCDF	105	9
	2,3,7,8-TCDD	90	6
	1,2,3,7,8-PeCDD	100	4
	1,2,3,4,7,8-HxCDD	107	5
	1,2,3,6,7,8-HxCDD	89	4
	1,2,3,7,8,9-HxCDD	106	6
	1,2,3,4,6,7,8-HpCDD	101	2
	1,2,3,4,6,7,8,9-OCDD	92	8

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	93	1
	PCB-#52	86	2
	PCB-#101	95	1
	PCB-#153	90	3
	PCB-#138	98	2
	PCB-#180	96	2
	PCB-#81	97	1
	PCB-#77	96	1
	PCB-#126	110	5
	PCB-#169	102	4
	PCB-#123	100	2
	PCB-#118	99	3
	PCB-#114	97	4
	PCB-#105	98	5
	PCB-#167	96	1
	PCB-#156	100	4
	PCB-#157	103	4
	PCB-#189	100	2