

# Quality Control Certificate

**Product:** Universal Column

**Product No.:** 19511

**Lot No.:** 716448

**Storage Recommendations:** Store the column at room temperature below 25°C

**Description:** The universal column is part of a 3- or 4-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

## Quality Control Release Inspection and Test Specification

**Test Procedure:** A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated via DEva and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

**Results Blank Value:**

PCDD/F-TEQ:	0,06	pg/column
	(crit: <	0,7 pg/column)
dl-PCB-TEQ:	0,004	pg/column
	(crit: <	0,05 pg/column)
Sum Indikator PCB:	8,9	pg/column
	(crit: <	100 pg/column)

**Results Recoveries:**

PCDD/F	75	to	105	%	(crit: 70	to	120)
PCB	71	to	106	%	(crit: 70	to	120)

This is to certify that universal column, Lot 716448, passed the required test specifications and is released for sale.

date: 18.02.2022 sign.: \_\_\_\_\_

*T. Kehmeier*

The company LCTech GmbH is certified according to ISO 9001:2015



**Hazards:** NOT FOR HUMAN OR DRUG USE!

The universal column is designed and prepared for usage with the alumina/florisil column and carbon column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.

**Quality Control:** All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.

**Quality Management:** This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)

**Documentation / Data Attached:** Table 1 & 2: Blank values of PCDD/F and PCB  
Table 3 & 4: 13C-Recoveries of PCDD/F and PCB

**Analytcs:** All the columns (n>5) have to perform a clean-up of a solvent blank (10 mL n-hexane), spiked with a 13C - labelled quantifier-standard solution with a single column method onto a DEXTech Plus system. The fractions 1 (PCB) and 2 (PCDD/F) are spiked with 13C - labelled recovery- standard solutions and evaporated with the D-EVA vacuum centrifuge. The extracts are measured with a HRMS-DFS from Thermo Fisher Scientific with a resolution of R > 10000. The HRGCs are equipped with 60 m DB5 MS columns. For PCDD/F 5µL are injected via PTV, for PCB 2µL via SSL.

**Remarks:** Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of silica or filters without any effect on the clean-up.

*The company LCTech GmbH is certified according to ISO 9001:2015*



**Results:**

Lockmass check:

No significant disturbances, or indicators for contaminations are detected.

Blanks:

Table 1: PCDD/F blank (n=5)

Table 2: PCB blank (n=5)

Congeneres:	[pg/column]:
2,3,7,8-TCDF	<0,036
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<0,081
1,2,3,4,7,8-HxCDF	<0,027
1,2,3,6,7,8-HxCDF	<dl
2,3,4,6,7,8-HxCDF	<0,045
1,2,3,7,8,9-HxCDF	<0,045
1,2,3,4,6,7,8-HpCDF	<0,063
1,2,3,4,7,8,9-HpCDF	<dl
OCDF	<0,054
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	<0,054
1,2,3,4,7,8-HxCDD	<0,027
1,2,3,6,7,8-HxCDD	<0,108
1,2,3,7,8,9-HxCDD	<0,027
1,2,3,4,6,7,8-HpCDD	<0,09
OCDD	0,34

Congeneres:	[pg/column]:
PCB 28	2
PCB 52	2,65
PCB 77	0,09
PCB 81	0,029
PCB 101	1,67
PCB 123	0,1901
PCB 118	0,57
PCB 114	0,337
PCB 105	0,44
PCB 126	0,0379
PCB 153	0,84
PCB 138	0,99
PCB 167	0,447
PCB 156	0,29
PCB 157	0,147
PCB 169	<0,027
PCB 180	0,29
PCB 189	0,282

TEQ (WHO 2005)	
lower bound	0,06
upper bound	0,07

TEQ (WHO 2005)	
lower bound	0,0044
upper bound	0,0044

Sum DIN PCB	8,9
-------------	-----

The company LCTech GmbH is certified according to ISO 9001:2015

**Results:**

13C-Recoveries

Table 3: PCDD/F 13C-recoveries (n=5)

Congeneres:	13C rec [%]
2,3,7,8-TCDF	86
1,2,3,7,8-PeCDF	88
2,3,4,7,8-PeCDF	89
1,2,3,4,7,8-HxCDF	85
1,2,3,6,7,8-HxCDF	87
2,3,4,6,7,8-HxCDF	93
1,2,3,7,8,9-HxCDF	93
1,2,3,4,6,7,8-HpCDF	98
1,2,3,4,7,8,9-HpCDF	105
OCDF	102
2,3,7,8-TCDD	81
1,2,3,7,8-PeCDD	87
1,2,3,4,7,8-HxCDD	104
1,2,3,6,7,8-HxCDD	75
1,2,3,7,8,9-HxCDD	94
1,2,3,4,6,7,8-HpCDD	101
OCDD	99

Table 4: PCB 13C-recoveries (n=5)

Congeneres:	13C rec [%]
PCB 28	95
PCB 52	101
PCB 77	71
PCB 81	71
PCB 101	103
PCB 123	93
PCB 118	83
PCB 114	98
PCB 105	87
PCB 126	73
PCB 153	105
PCB 138	106
PCB 167	86
PCB 156	87
PCB 157	81
PCB 169	71
PCB 180	101
PCB 189	79

The company LCTech GmbH is certified according to ISO 9001:2015

