

Quality Control Certificate

Product: **Universal Column**

Product No.: **19511**

Lot No.: **722976**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Universal Column is part of a 3- or 4-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of $R > 10000$.

Results Blank Value:	PCDD/F-TEQ:	0,07	pg/column						
		(crit: <	0,70	pg/column)					
	dl-PCB-TEQ:	0,003	pg/column						
		(crit: <	0,05	pg/column)					
	Sum Total PCB:	17,2	pg/column						
		(crit: <	300	pg/column)					
Results Recoveries:	PCDD/F	85	to	119	%	(crit: 70	to	120	%)
	PCB	80	to	112	%	(crit: 70	to	120	%)

This is to certify that the Universal Column, Lot 722976, passed the required test specifications and is released for sale.

date: 09.01.2026 sign.: M.Bradis

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 19511 - 722976

Hazards: NOT FOR HUMAN OR DRUG USE!

The Universal Column is designed and prepared for usage with the Alumina/Florisil Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.

Quality Control: All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.

Quality Management: This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)

Documentation /
Data Attached: table 1 & 2: blankvalues of PCDD/F and PCB
table 3 & 4: 13C-Recoveries of PCDD/F and PCB

Analytics This is to certify that the Universal Column, Lot , passed the required test specifications and is released for sale.

Remarks Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of silica or filters without any effect on the clean-up.

QC-Certificate - 19511 - 722976

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

sample amount [pg/column]	[pg/column]
2,3,7,8-TCDF	<0,036
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	<0,018
2,3,4,6,7,8-HxCDF	<dl
1,2,3,7,8,9-HxCDF	<dl
1,2,3,4,6,7,8-HpCDF	<0,063
1,2,3,4,7,8,9-HpCDF	<dl
1,2,3,4,6,7,8,9-OCDF	<dl
2,3,7,8-TCDD	<0,036
1,2,3,7,8-PeCDD	<dl
1,2,3,4,7,8-HxCDD	<0,027
1,2,3,6,7,8-HxCDD	<0,108
1,2,3,7,8,9-HxCDD	<dl
1,2,3,4,6,7,8-HpCDD	<0,09
1,2,3,4,6,7,8,9-OCDD	0,56

Table 2: PCB blank

sample amount [pg/sample]	[pg/column]
PCB-#28	6,64
PCB-#52	7,02
PCB-#101	1,85
PCB-#153	0,96
PCB-#138	0,71
PCB-#180	<0,18
PCB-#81	0,05
PCB-#77	0,18
PCB-#126	0,0247
PCB-#169	<0,027
PCB-#123	0,03
PCB-#118	0,67
PCB-#114	0,012
PCB-#105	<0,081
PCB-#167	0,027
PCB-#156	<dl
PCB-#157	<0,018
PCB-#189	0,035

PCDD/F TEQ (2005) [pg/column]
lower bound 0,03
upper bound 0,07

PCB-TEQ [pg/column]
lower bound 0,003
upper bound 0,003
Sum DIN 17,2

Table 3: PCDD/F recoveries

PCDD/F 13C Recoveries [%]	[%]	RSD [%]
2,3,7,8-TCDF	91	3
1,2,3,7,8-PeCDF	95	3
2,3,4,7,8-PeCDF	99	2
1,2,3,4,7,8-HxCDF	109	2
1,2,3,6,7,8-HxCDF	117	2
2,3,4,6,7,8-HxCDF	107	2
1,2,3,7,8,9-HxCDF	113	4
1,2,3,4,6,7,8-HpCDF	113	3
1,2,3,4,7,8,9-HpCDF	89	2
1,2,3,4,6,7,8,9-OCDF	92	3
2,3,7,8-TCDD	90	3
1,2,3,7,8-PeCDD	99	2
1,2,3,4,7,8-HxCDD	111	2
1,2,3,6,7,8-HxCDD	95	2
1,2,3,7,8,9-HxCDD	119	4
1,2,3,4,6,7,8-HpCDD	101	2
1,2,3,4,6,7,8,9-OCDD	85	4

Table 4: PCB recoveries

PCB 13C Recoveries [%]	[%]	RSD [%]
PCB-#28	85	3
PCB-#52	93	2
PCB-#101	92	2
PCB-#153	93	3
PCB-#138	93	1
PCB-#180	98	3
PCB-#81	101	2
PCB-#77	104	3
PCB-#126	105	3
PCB-#169	112	3
PCB-#123	83	7
PCB-#118	80	11
PCB-#114	87	4
PCB-#105	82	9
PCB-#167	82	10
PCB-#156	83	8
PCB-#157	84	11
PCB-#189	89	8