

## Quality Control Certificate

Product: **Standard Column**  
 Product No.: 19512  
 Lot No.: **722870**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The standard Column is part of a 3- or 4-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.


### Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,06	pg/column
		(crit: <	0,70 pg/column)
	dl-PCB-TEQ:	0,0017	pg/column
	(crit: <	0,05 pg/column)	
	Sum Total PCB:	46,2	pg/column
	(crit: <	300 pg/column)	

Results Recoveries:	PCDD/F	81	to	120	%	(crit: 70	to	120	%)
	PCB	83	to	104	%	(crit: 70	to	120	%)

This is to certify that the Standard Column, Lot 722870, passed the required test specifications and is released for sale.

date: 12.05.2026 sign.: 

The company LCTech GmbH is certified according to ISO 9001



## QC-Certificate - 19512 - 722870

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The standard Column is designed and prepared for usage with the Alumina/Florisol Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 &amp; 2: blankvalues of PCDD/F and PCB table 3 &amp; 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the standard Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of silica or filters without any effect on the clean-up.</p>

QC-Certificate - 19512 - 722870

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	<dl
2,3,4,6,7,8-HxCDF	<dl
1,2,3,7,8,9-HxCDF	<dl
1,2,3,4,6,7,8-HpCDF	<dl
1,2,3,4,7,8,9-HpCDF	<dl
1,2,3,4,6,7,8,9-OCDF	<dl
2,3,7,8-TCDD	<0,036
1,2,3,7,8-PeCDD	<dl
1,2,3,4,7,8-HxCDD	<dl
1,2,3,6,7,8-HxCDD	<dl
1,2,3,7,8,9-HxCDD	<dl
1,2,3,4,6,7,8-HpCDD	<0,09
1,2,3,4,6,7,8,9-OCDD	0,26

Table 2: PCB blank

	[pg/column]
PCB-#28	29,2
PCB-#52	14,25
PCB-#101	1,76
PCB-#153	0,52
PCB-#138	0,44
PCB-#180	<0,18
PCB-#81	0,03
PCB-#77	0,225
PCB-#126	0,0163
PCB-#169	<dl
PCB-#123	0,02
PCB-#118	0,64
PCB-#114	0,009
PCB-#105	0,16
PCB-#167	<dl
PCB-#156	<dl
PCB-#157	0,03
PCB-#189	<0,0072

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,02
upper bound	0,06

PCB-TEQ	[pg/column]
lower bound	0,0017
upper bound	0,002
Sum DIN	46,2

QC-Certificate - 19512 - 722870

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	101	2
	1,2,3,7,8-PeCDF	98	4
	2,3,4,7,8-PeCDF	99	2
	1,2,3,4,7,8-HxCDF	113	6
	1,2,3,6,7,8-HxCDF	119	4
	2,3,4,6,7,8-HxCDF	120	4
	1,2,3,7,8,9-HxCDF	120	3
	1,2,3,4,6,7,8-HpCDF	116	2
	1,2,3,4,7,8,9-HpCDF	88	3
	1,2,3,4,6,7,8,9-OCDF	91	4
	2,3,7,8-TCDD	96	3
	1,2,3,7,8-PeCDD	98	5
	1,2,3,4,7,8-HxCDD	119	4
	1,2,3,6,7,8-HxCDD	104	3
	1,2,3,7,8,9-HxCDD	119	3
	1,2,3,4,6,7,8-HpCDD	97	2
	1,2,3,4,6,7,8,9-OCDD	81	2

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	90	2
	PCB-#52	99	5
	PCB-#101	90	1
	PCB-#153	91	2
	PCB-#138	94	2
	PCB-#180	90	3
	PCB-#81	101	3
	PCB-#77	104	2
	PCB-#126	94	4
	PCB-#169	90	5
	PCB-#123	87	3
	PCB-#118	87	4
	PCB-#114	91	4
	PCB-#105	86	4
	PCB-#167	83	2
	PCB-#156	85	2
	PCB-#157	84	2
	PCB-#189	83	3