

Quality Control Certificate

Product: Smart Column

Product No.: 19513

Lot No.:

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Smart Column is part of a 3-column setup for the sample cleanup of environmental-, food- / feed- and similar matrices. It is designed for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners with the DEXTech systems from LCTech GmbH

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated via D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,37	pg/column
		(crit: <	0,7 pg/column)
	dl-PCB-TEQ:	0,033	pg/column
	(crit: <	0,05 pg/column)	
	Sum Indikator PCB:	33,53	pg/column
	(crit: <	100 pg/column)	

Results Recoveries:	PCDD/F	88	to	111	%	(crit: 70	to	120)
	PCB	79	to	97	%	(crit: 70	to	120)

This is to certify that the 19513, Lot , passed the required test specifications and is released for sale.

date: 07.06.2022 sign.: _____

T. Kehmeier

The company LCTech GmbH is certified according to ISO 9001:2015



Hazards: NOT FOR HUMAN OR DRUG USE!

The Smart Column is designed and prepared for usage with the Alumina/Florisil and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.

Quality Control: All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.

Quality Management: This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)

Documentation / Data Attached: Table 1 & 2: Blank values of PCDD/F and PCB
Table 3 & 4: 13C-Recoveries of PCDD/F and PCB

Analytcs: All the columns (n>5) have to perform a clean-up of a solvent blank (10 mL n-hexane), spiked with a 13C - labelled quantifier-standard solution with a single column method onto a DEXTech Plus system. The fractions 1 (PCB) and 2 (PCDD/F) are spiked with 13C - labelled recovery- standard solutions and evaporated with the D-EVA vacuum centrifuge. The extracts are measured with a HRMS-DFS from Thermo Fisher Scientific with a resolution of R > 10000. The HRGCs are equipped with 60 m DB5 MS columns. For PCDD/F 5µL are injected via PTV, for PCB 2µL via SSL.

Remarks: Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of silica or filters without any effect on the clean-up.

The company LCTech GmbH is certified according to ISO 9001:2015



Results:

Lockmass check:

No significant disturbances, or indicators for contaminations are detected.

Blanks:

Table 1: PCDD/F blank (n=5)

Table 2: PCB blank (n=5)

Congeneres: [pg/column]:

2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	0,13
2,3,4,7,8-PeCDF	<0,081
1,2,3,4,7,8-HxCDF	0,211
1,2,3,6,7,8-HxCDF	0,179
2,3,4,6,7,8-HxCDF	0,22
1,2,3,7,8,9-HxCDF	0,19
1,2,3,4,6,7,8-HpCDF	0,4
1,2,3,4,7,8,9-HpCDF	0,344
OCDF	0,22
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	0,14
1,2,3,4,7,8-HxCDD	0,237
1,2,3,6,7,8-HxCDD	0,71
1,2,3,7,8,9-HxCDD	0,29
1,2,3,4,6,7,8-HpCDD	0,38
OCDD	0,91

TEQ (WHO 2005)	
lower bound	0,37
upper bound	0,38

Congeneres: [pg/column]:

PCB 28	2,22
PCB 52	5,39
PCB 77	0,36
PCB 81	0,189
PCB 101	5,28
PCB 123	0,3775
PCB 118	1,13
PCB 114	0,4695
PCB 105	0,83
PCB 126	0,2265
PCB 153	5,5
PCB 138	8,37
PCB 167	0,595
PCB 156	0,58
PCB 157	0,425
PCB 169	0,321
PCB 180	6,17
PCB 189	0,952

TEQ (WHO 2005)	
lower bound	0,0325
upper bound	0,0325

Sum DIN PCB	33,53
-------------	-------

The company LCTech GmbH is certified according to ISO 9001:2015



Results:

13C-Recoveries

Table 3: PCDD/F 13C-recoveries (n=5)

Congeneres:	13C rec [%]
2,3,7,8-TCDF	100
1,2,3,7,8-PeCDF	97
2,3,4,7,8-PeCDF	91
1,2,3,4,7,8-HxCDF	90
1,2,3,6,7,8-HxCDF	97
2,3,4,6,7,8-HxCDF	98
1,2,3,7,8,9-HxCDF	101
1,2,3,4,6,7,8-HpCDF	107
1,2,3,4,7,8,9-HpCDF	96
OCDF	111
2,3,7,8-TCDD	96
1,2,3,7,8-PeCDD	91
1,2,3,4,7,8-HxCDD	102
1,2,3,6,7,8-HxCDD	88
1,2,3,7,8,9-HxCDD	100
1,2,3,4,6,7,8-HpCDD	105
OCDD	105

Table 4: PCB 13C-recoveries (n=5)

Congeneres:	13C rec [%]
PCB 28	95
PCB 52	92
PCB 77	91
PCB 81	91
PCB 101	97
PCB 123	90
PCB 118	86
PCB 114	94
PCB 105	90
PCB 126	86
PCB 153	90
PCB 138	93
PCB 167	79
PCB 156	91
PCB 157	90
PCB 169	90
PCB 180	96
PCB 189	80

The company LCTech GmbH is certified according to ISO 9001:2015

