

Quality Control Certificate

Product: **EVOLUTION Aloxx Column**
Product No.: 20087
Lot No.: **724032**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The EVOLUTION Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of $R > 10000$.

Results Blank Value:	PCDD/F-TEQ:	0,05	pg/column
		(crit: <	0,70 pg/column)
	dl-PCB-TEQ:	0,0022	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	48	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	84	to	115	%	(crit: 70	to	120	%)
	PCB	95	to	108	%	(crit: 70	to	120	%)

This is to certify that the EVOLUTION Aloxx Column, Lot 724032, passed the required test specifications and is released for sale.

date: 09.07.2026 sign.: 

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 20087 - 724032

Hazards:	NOT FOR HUMAN OR DRUG USE! The Alumina Column is designed and prepared for usage with the Universal/standard & Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.
Quality Control:	All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.
Quality Management:	This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)
Documentation / Data Attached:	table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB
Analytics	This is to certify that the EVOLUTION Alumina Column, Lot , passed the required test specifications and is released for sale.
Remarks	n/a

QC-Certificate - 20087 - 724032

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
sample amount [pg/column]	
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	<dl
2,3,4,6,7,8-HxCDF	<dl
1,2,3,7,8,9-HxCDF	<dl
1,2,3,4,6,7,8-HpCDF	<dl
1,2,3,4,7,8,9-HpCDF	<dl
1,2,3,4,6,7,8,9-OCDF	<dl
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	<dl
1,2,3,4,7,8-HxCDD	<dl
1,2,3,6,7,8-HxCDD	<dl
1,2,3,7,8,9-HxCDD	<dl
1,2,3,4,6,7,8-HpCDD	<0,09
1,2,3,4,6,7,8,9-OCDD	0,35

Table 2: PCB blank

	[pg/column]
sample amount [pg/sample]	
PCB-#28	27,56
PCB-#52	16,44
PCB-#101	2,23
PCB-#153	0,78
PCB-#138	0,96
PCB-#180	<0,18
PCB-#81	0,04
PCB-#77	0,145
PCB-#126	<dl
PCB-#169	0,069
PCB-#123	0,03
PCB-#118	0,89
PCB-#114	0,028
PCB-#105	0,24
PCB-#167	0,065
PCB-#156	<0,126
PCB-#157	0,05
PCB-#189	0,029

PCDD/F TEQ (2005)	[pg/column]
lower bound	0
upper bound	0,05

PCB-TEQ	[pg/column]
lower bound	0,0022
upper bound	0,0023
Sum DIN	48



QC-Certificate - 20087 - 724032

Table 3: PCDD/F recoveries

		RSD [%]	
		[%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	91	7
	1,2,3,7,8-PeCDF	102	6
	2,3,4,7,8-PeCDF	98	7
	1,2,3,4,7,8-HxCDF	107	7
	1,2,3,6,7,8-HxCDF	114	8
	2,3,4,6,7,8-HxCDF	110	6
	1,2,3,7,8,9-HxCDF	113	5
	1,2,3,4,6,7,8-HpCDF	109	5
	1,2,3,4,7,8,9-HpCDF	103	5
	1,2,3,4,6,7,8,9-OCDF	101	5
	2,3,7,8-TCDD	84	7
	1,2,3,7,8-PeCDD	101	5
	1,2,3,4,7,8-HxCDD	112	4
	1,2,3,6,7,8-HxCDD	97	6
	1,2,3,7,8,9-HxCDD	115	6
	1,2,3,4,6,7,8-HpCDD	105	6
	1,2,3,4,6,7,8,9-OCDD	91	5

Table 4: PCB recoveries

		RSD [%]	
		[%]	
PCB 13C Recoveries [%]	PCB-#28	100	4
	PCB-#52	98	5
	PCB-#101	96	3
	PCB-#153	97	3
	PCB-#138	97	3
	PCB-#180	97	2
	PCB-#81	97	3
	PCB-#77	101	5
	PCB-#126	103	4
	PCB-#169	95	3
	PCB-#123	108	4
	PCB-#118	102	4
	PCB-#114	101	4
	PCB-#105	103	4
	PCB-#167	98	3
	PCB-#156	99	3
	PCB-#157	101	3
	PCB-#189	95	2